



Basic Introduction:

According to: IEC 60065:9.1.3, GB 8898:9.1.3, EN 61010-1, UL 3101-1, and CSA 1010-1.
Openings of the enclosure

Technical parameters:

1. Rod length: 100 mm
2. Rod diameter: 4 mm
3. Circular-baffle diameter: 35 mm
4. Handle diameter: 10 mm
5. Handle length: 100 mm

Application:

The apparatus shall be so designed that suspended foreign bodies cannot become HAZARDOUS LIVE, when introduced through ventilation or other holes.

Compliance is checked by applying to the holes a metal test pin having a diameter of 4 mm and a length of 100 mm. The test pin is suspended freely from one end, the penetration is limited to the length of the test pin.

Note : The test pin shall not become HAZARDOUS LIVE.

