



**IEC61032 Test Probe 18, Children Test Finger Probe**

This probe is intended to simulate access to hazardous parts by children of more than 36 months and less than 14 years.

**Conforms to:**

Standard IEC61032-1997, IEC60529-2001, GB/T4208-2008 and UL etc.

**Technology Parameter:**

Knurled finger diameter	8.6mm
Head radius	4.3mm
Knurled finger length	57.9mm (three knurls, 19.3mm/knurl)

**Notes:**

The extension of the handle represents the arm of the child. And the handle is provided with an extension 451.6mm long, and the probe should be applied with or without this extension, whichever is the more onerous condition.

Both joints shall permit movement in the same plane and the same direction through an angle of 90o..

f)

