



Test Probe 11 of IEC 61032

The **Rigid Test Finger** is a precision test probe made according to **Figure 7 (Fig. 7) of the IEC 61032 (Test probe 11)** and is used to simulate a human finger. It is also used by the standards of CSA, IRAM, UL. IEC 60335, IRAM 4220-1 and in most of the rules involved in the verification of accessibility to live parts.

Specifications	
Material:	Stainless Steel.
Handle:	Nylon
According to:	IEC 61032. IEC 60335-1. IEC 60601.



