



**Purpose:**

This Rigid Test Finger may be used to verify the protection of persons against access to hazardous parts and to verify the mechanical strength of openings in the enclosure or internal barriers.

**Conforms to:**

Standard IEC 61032-1997 Figure 7 (Test Probe 11), IEC 60335, IRAM 4220-1, and UL etc.

**Technical Parameter:**

Test probe diameter: 12 mm

Test probe length: 80 mm

Baffle plate thickness: 5 mm

Baffle plate diameter: 50 mm

